

DOCKET NO. SA 516

EXHIBIT NO. 8M

NATIONAL TRANSPORTATION SAFETY BOARD

WASHINGTON, D.C.

**CERTIFICATE OF CALIBRATION FOR DIGITAL CONDUCTIVITY
METER SN 14912**

by

Gordon J. Hookey



emcee electronics, inc.

520 CYPRESS AVENUE • VENICE, FLORIDA 34292

CERTIFICATE OF CALIBRATION

DIGITAL CONDUCTIVITY METER

MODEL 1152

P/N 152-00-000 0

S/N 14912

J/N 105933

- | | |
|--|-----------|
| | PASS |
| 1. ELECTRICAL CALIBRATION | <u>✓</u> |
| 2. HUMIDITY EXPOSURE | <u>NA</u> |
| 3. VISUAL INSPECTION | <u>✓</u> |
| 4. FUEL EVALUATION VS. PRECISION METER | |


| LOW | | HIGH | |
|-----------|------------------|------------|--------------------|
| PRECISION | READING | PRECISION | READING |
| <u>67</u> | <u>68</u> ± 5 | <u>299</u> | <u>300</u> ± 20 |

5. PROBE NUMBER 41

DATE 5/6/96

Q.C. TECHNICIAN K. Van Schoick

DATE 5/6/96

VERIFIED BY: P. Kon 

ATTACHMENT